



# INSTRUCTIONS FOR DISPOSABLE INFUSION SETS

## INDICATIONS FOR USE

The product is gravity feed only and used to infusion.

### Main structure and properties

- Tubing and drip chamber are transparent, can easy to observe;
- Tubing is flexible enough to resist winding and twisting;
- The design of closure-piercing device to make it convenient for handle and usage;
- Flow regulator can make the flow rate change from minimum to maximum.

### USAGE

- Firstly, check valid period and completeness of package before using;
- Tear the package at the cut part and take out the infusion set;
- Close the regulator, pull out protective cap of closure-piercing device and push the spike into the disinfection closure of fluid container;
- Put fluid container upside down, press chamber and half of it was filled with injection solution; then open the regulator to expel air in tube.
- Penetrate vein according aseptic manipulation after expelling air then fix it.
- Observing and then change flow rate to normal.
- Infusion operation must have the professionals nursing to implement and guardian.
- 20 drops=(1±0.1)ml[(1±0.1)g]

### PRECAUTIONS

- It may causes hypothermia-esp, weak patients;
- It may causes abnormal thythem of ventrictkes of heart because of fast flow rate;
- There should maintain the guardianship during the infusion in case of any accident;
- It should be sterile and free from pyrogen before package opened;.
- It has a valid period of five years, do not use if it is beyond valid date.
- Discard after singe use.

### WARNINGS

- Do not use if the package is opened or damaged, or protective cap is aparted, any part is missed, and foreign matter is found;
- It should be used by special trained person , and the skin should be correctly sterilized.
- The products can not be reused. The repetition use may cause cross infection, even serious illnesses such as hepatitis, HIV and other infectious disease.
- Always dispose of blood contaminated products in a manner consistent with established biohazard procedures.

### STORAGE

Product should be kept in a condition that relative humidity is less than 80%, it also should be free from corrosive gas, cool, dry with good ventilation and clean.

**Any serious incident that has occurred in relation to the device should be reported to the manufacturer and the competent authority of the Member State in which the user and/or patient is established.**

Unique device identifier	Batch code	Catalogue number	Caution	Sterilized using ethylene oxide	
Date of manufacture	Manufacturer	Consult Instructions For Use	Do not re-use	Do not use if package is damaged	
Medical device	Use-by date	Do not re-sterilize	Fragile, handle with care	Temperature limit	2862
Keep away from sunlight	Keep dry	Non-pyrogenic	This way up	Humidity limitation	
Single sterile barrier system	Sterile	Contains or presence of Phthalate		REV.: C/0 EFFECTIVE DATE: 2025-05-30	

Anhui Tiankang Medical Technology Co., Ltd.  
No. 228, Weiyi Road, Economic Development  
Zone, Tianchang City, Anhui Province, P.R. China.

Shanghai International Holding Corp. GmbH(Europe)  
Eiffelstrasse 80 20537 Hamburg, Germany  
Tel:0049-40-2513175